



SE-CHECK™

HEMATOLOGY CONTROL FOR SYSMEX SE-SERIES ANALYZERS OPEN MODE ASSAY

Expiration Date: 2009-01-11		Quality Control Data Due Date #1: 11-Dec-08			Quality Control Data Due Date #2: 11-Jan-09		
Lot Number:		83010510		83010511		83010512	
Control:		Low Abnormal		Normal		High Abnormal	
PARAMETERS		MEAN	EXPECTED RANGE	MEAN	EXPECTED RANGE	MEAN	EXPECTED RANGE
O P E N	WBC (10 ⁹ /L)	3.10	2.70 - 3.51	7.00	6.31 - 7.69	16.87	15.20 - 18.54
	RBC (10 ¹² /L)	2.44	2.31 - 2.57	4.30	4.17 - 4.43	5.27	5.11 - 5.43
	HGB (g/dL)	6.5	6.2 - 6.8	13.0	12.7 - 13.3	17.4	16.8 - 18.0
	HCT (%)	19.6	18.3 - 21.0	38.1	36.0 - 40.2	50.3	47.6 - 53.0
	MCV (fL)	80.5	71.1 - 91.0	88.6	81.1 - 96.5	95.5	87.6 - 103.8
	MCH (pg)	26.5	24.0 - 29.4	30.2	28.6 - 32.0	33.0	30.9 - 35.3
	MCHC (g/dL)	33.0	29.4 - 37.1	34.1	31.5 - 37.1	34.6	31.6 - 37.9
	RDW-SD (fL)	44.3	41.7 - 46.9	42.9	39.9 - 46.0	46.9	43.4 - 50.5
	RDW-CV (%)	15.2	14.8 - 15.7	13.6	12.8 - 14.3	14.5	13.6 - 15.3
	PLT (10 ⁹ /L)	51	42 - 61	218	204 - 231	524	495 - 553
	MPV (fL)	9.4	8.6 - 10.1	9.7	9.0 - 10.3	9.8	9.2 - 10.5
	NEUT# (10 ⁹ /L)	2.19	1.81 - 2.56	4.95	4.10 - 5.79	12.09	10.47 - 13.72
	LYMPH# (10 ⁹ /L)	0.71	0.56 - 0.87	1.61	1.16 - 2.06	3.74	2.69 - 4.79
	MONO# (10 ⁹ /L)	0.20	0.08 - 0.32	0.44	0.14 - 0.75	1.04	0.45 - 1.63
	EO# (10 ⁹ /L)	2.38	1.97 - 2.79	5.25	4.54 - 5.96	12.62	10.44 - 14.81
	BASO# (10 ⁹ /L)	3.10	2.56 - 3.64	6.63	5.82 - 7.44	14.93	12.97 - 16.88
NEUT%	70.5	64.2 - 76.8	70.7	63.5 - 77.8	71.7	65.6 - 77.8	
LYMPH%	23.0	17.4 - 28.6	23.0	17.2 - 28.9	22.2	15.7 - 28.6	
MONO%	6.5	2.7 - 10.4	6.3	1.9 - 10.8	6.2	2.3 - 10.0	
EO%	76.6	60.4 - 92.9	75.0	61.6 - 88.4	74.8	62.1 - 87.5	
BASO%	100.1	80.0 - 120.2	94.7	79.7 - 109.7	88.5	75.4 - 101.6	

Sysmex recommends that laboratories establish their own Q.C. Target Values for each new lot number by collecting at least 10 data points per control level over 5 days. The mean Q.C. Target Values from this initial data collection should be inside the above Expected Ranges. These Expected Ranges represent estimates of interlaboratory variation, and are not intended for use as the laboratory's internal Q.C. file limits. Sysmex recommends that each laboratory establish its own Q.C. file limits based on the laboratory's historical coefficient of variation. See your Sysmex Training Manual for more information on establishing Q.C. file limits.

Distributed by:
Sysmex America, Inc.
Mundelein, IL

350093
Rev 3, 17-Nov-06
Nov-08



SE-CHECK™

HEMATOLOGY CONTROL FOR SYSMEX SE-SERIES ANALYZERS CLOSED MODE ASSAY

Expiration Date: 2009-01-11		Quality Control Data Due Date #1: 11-Dec-08			Quality Control Data Due Date #2: 11-Jan-09		
Lot Number:		83010510		83010511		83010512	
Control:		Low Abnormal		Normal		High Abnormal	
PARAMETERS		MEAN	EXPECTED RANGE	MEAN	EXPECTED RANGE	MEAN	EXPECTED RANGE
CLOSED	WBC (10 ⁹ /L)	2.98	2.64 - 3.32	6.72	6.06 - 7.38	15.77	14.21 - 17.34
	RBC (10 ¹² /L)	2.43	2.34 - 2.52	4.34	4.18 - 4.51	5.21	5.01 - 5.42
	HGB (g/dL)	6.5	6.2 - 6.7	13.1	12.5 - 13.6	17.3	16.6 - 18.0
	HCT (%)	19.5	18.4 - 20.7	38.5	36.1 - 40.8	49.8	46.8 - 52.8
	MCV (fL)	80.5	73.0 - 88.5	88.6	80.1 - 97.7	95.5	86.4 - 105.4
	MCH (pg)	26.7	24.7 - 28.8	30.1	27.8 - 32.6	33.2	30.6 - 36.0
	MCHC (g/dL)	33.1	30.1 - 36.5	34.0	30.7 - 37.6	34.8	31.4 - 38.5
	RDW-SD (fL)	44.3	42.0 - 46.7	42.9	40.6 - 45.2	46.9	44.0 - 49.9
	RDW-CV (%)	15.2	14.6 - 15.9	13.6	13.0 - 14.1	14.5	13.8 - 15.2
	PLT (10 ⁹ /L)	49	41 - 56	211	187 - 235	493	439 - 547
	MPV (fL)	9.4	8.8 - 9.9	9.7	9.2 - 10.1	9.8	9.3 - 10.4
	NEUT# (10 ⁹ /L)	2.10	1.77 - 2.43	4.76	3.97 - 5.55	11.32	9.45 - 13.20
	LYMPH# (10 ⁹ /L)	0.68	0.55 - 0.82	1.55	1.12 - 1.97	3.50	2.55 - 4.44
	MONO# (10 ⁹ /L)	0.19	0.06 - 0.32	0.42	0.12 - 0.71	0.95	0.22 - 1.68
	EO# (10 ⁹ /L)	2.28	1.81 - 2.75	5.04	4.03 - 6.05	11.80	9.37 - 14.23
	BASO# (10 ⁹ /L)	3.01	2.35 - 3.67	6.49	5.03 - 7.95	14.37	10.67 - 18.08
	NEUT%	70.6	63.6 - 77.6	70.8	61.9 - 79.6	71.8	65.4 - 78.2
LYMPH%	23.0	18.5 - 27.5	23.0	16.5 - 29.6	22.2	16.2 - 28.2	
MONO%	6.4	2.1 - 10.7	6.2	2.5 - 9.9	6.0	2.2 - 9.9	
EO%	76.6	60.3 - 93.0	75.0	60.1 - 89.8	74.8	59.0 - 90.6	
BASO%	101.1	76.4 - 125.9	96.6	74.5 - 118.7	91.1	64.3 - 118.0	

Sysmex recommends that laboratories establish their own Q.C. Target Values for each new lot number by collecting at least 10 data points per control level over 5 days. The mean Q.C. Target Values from this initial data collection should be inside the above Expected Ranges. These Expected Ranges represent estimates of interlaboratory variation, and are not intended for use as the laboratory's internal Q.C. file limits. Sysmex recommends that each laboratory establish its own Q.C. file limits based on the laboratory's historical coefficient of variation. See your Sysmex Training Manual for more information on establishing Q.C. file limits.

Distributed by:
Sysmex America, Inc.
Mundelein, IL

350093
Rev 3, 17-Nov-06
Nov-08